

TEL: (662) 318 9442 - 4
FAX: (662) 318 3268
EMAIL: info@warubbermate.co.th
WEBSITE: www.warubbermate.co.th



W.A. RUBBERMATE CO., LTD



4 SOI RAMKHAMHAENG 19 (CHAREONPLOY)
RAMKHAMHAENG ROAD, HUAMARK
BANGKAPI, BANGKOK 10240, THAILAND

CE EU DECLARATION OF CONFORMITY CE

According to the Medical Device Directive 93/42/EEC

Manufacturer's Name : W.A. RUBBERMATE CO., LTD
Manufacturer's Address : 4 SOI RAMKHAMHAENG 19 (CHAREONPLOY),
RAMKHAMHAENG ROAD, HUAMARK, BANGKAPI,
BANGKOK 10240, THAILAND

Quality Management System : ISO 9001:2008, ISO 13485: 2003

Name of Device : Nitrile Examination Gloves
Type : Powder Free
Classification : Class I, Non Sterile
Conformity Route : Annex VII

We herewith declare in our own responsibility that the above-mentioned product(s) with CE mark meet the provisions of the EC Council Directive 93/42/EEC 14th June 1993 concerned medical devices, amended by Council Directive 2007/47/EC. All supporting documentations are retained under the premise of the manufacturer (Head of Quality department).

General applicable directives:
Council Directive 93/42/EEC 14th June 1993 concerned medical devices, amended by Council Directive 2007/47/EC.


For evaluation of the compliance with these Directives, the following standards were applied:

- EN-455 – Parts I, II, III
- EN-374 – Parts I, II, III
- ISO 2859-1-99
- ASTM D6319-99

Name of Signatory: Benchamad Kongwaree
Title: Managing Director
Date: 31 January 2013

Signature:

Stamp of the Manufacturer





W.A. Rubbermate Co., Ltd.
WATTANACHAI RUBBERMATE Co., Ltd.
www.warubbermate.co.th AWARDED ISO9001 & ISO13485

NITRILE EXAMINATION GLOVES

This Product Does Not Contain Any Natural Rubber Latex

MATERIAL SAFETY DATA SHEET

IDENTIFICATION

Product Name : W.A. Rubbermate Co., Ltd
Product : Nitrile Examination Gloves
Lightly Powdered and Powder Free

Case # : Not Established
IATO / IACO : Not Restricted
Hazard Rating : Not Restricted
D.O.T. Rating : Not Restricted

PHYSICAL DATA

Nitrile Butadiene Rubber : 95.8%
Sulfur : 1%
Zinc Oxide : 1.2%
Accelerators : 1%
Others : 1%
Donning Agent : Food Grade Absorbable Dusting Powder USP Cornstarch
(Powdered Nitrile Examination Gloves Only)

FIRE AND EXPLOSION

Stability : Stable
Flash Point : N.A.
Extinguishing Media : Foam, Carbon Dioxide, Water

NORMAL PRECAUTIONS

Ventilation : Not necessary under normal conditions.
Equipment : Not necessary
Spill / Clean Up : N/A
Eyes : N/A
Storage : No special precautions necessary
Store at room temperature



Notified Body 0321

Issued to: **W.A. Rubbermate Co., Ltd**
67/1 Moo4, Bypass Road
T. Nongmaidaeng, Muang,
Chonburi 20000,
Thailand

SATRA Client: P1436

EC Type-Examination Certificate

Number 7642 Issue 1 Extension 1

This is to certify that the product group reference NBR-FT-35 comprising the following products

Product Reference	Description
NBR-FT-35	Five fingered Blue powder free Nitrile gloves
Sizes 7-10	

Classification:

EN 388:2003	Level	EN 374-2:2003	
Abrasion resistance	0	Air Leak	Pass
Blade cut resistance	0	Water Leak	Pass
Tear resistance	0		
Puncture resistance	0		

EN 374-3:2003

	Level
Sodium Hydroxide 40%	6
n-Heptane	0
Methanol	0

Technical reports

SATRA: SPC0221961/1406/1, CHM0229406/1443/EN, SPC0221961/1406/SMcD

has been subject to an EC Type-examination in accordance with Article 10 of the PPE Directive (89/686/EEC) and has been shown to satisfy the relevant provisions of this Directive for the complex category through:

- i Testing to the following standard: **EN 374-1:2003 (Excluding clause 5.3.2), EN388:2003, EN420:2003+A1:2009**
- ii Examination of the relevant technical documentation.

You are therefore licensed to mark the product(s) listed above in accordance with Article 13 of Directive (89/686/EEC) and any relevant amending Directives once you have drawn up an EC declaration of product conformity. Please note that:

1. Full details of the certification and product are contained in the manufacturer's technical file
2. This certificate is only valid if embossed with the text – "SATRA – European Notified Body 0321"
3. This certificate is issued subject to the conditions on the reverse side of this certificate
4. CE marking of production is also reliant on current compliance with Directive 89/686/EEC Article 11

Signed:  (T. Holly)

Signed:  Date: 20th February 2015

On behalf of SATRA



Has-Pet

L
LARGE

**MAVİ NİTRİL
MUAYENE ELDİVENİ**

Pudrasız / Powder Free

100
Adet/Pcs.

BU ÜRÜN YALNIZ KLİNİK KULLANIM İÇİN DİZAYN EDİLMİŞTİR. AMELİYATLARDA KULLANILMAZ! CAUTION: This product is not intended to be used in surgeries.

www.nhmedical.com.tr

www.nhmedical.com.tr

www.nhmedical.com.tr

Has-Pet

PUDRASIZ / POWDER FREE

KÜÇÜK BOY / LARGE
Small / Blue

www.nhmedical.com.tr

www.nhmedical.com.tr

NCIMEDICAL

Tıbbi Malzemeler ve Ekipmanlar

Ordu, Zeytin Mah. Çarşıbaşı Sokak, 11. Sokak No: 51
04520 00000 / 04520 00002 / Türkiye

www.nhmedical.com.tr

NCIMEDICAL

Tıbbi Malzemeler ve Ekipmanlar

Ordu, Zeytin Mah. Çarşıbaşı Sokak, 11. Sokak No: 51
04520 00000 / 04520 00002 / Türkiye

Has-Pet

NİTRİL ELDİVEN ANTIKARI

- 1- Üstte dikaralıdır, altına keskin kenarları pürüz kenarları ve ucuna uzatılmıştır.
- 2- Yırtılmaları ve yırtılmaya karşı koruyucu yapıştırıcıdır.
- 3- Sterilize edilmiş ambalajda sunulmaktadır.
- 4- Tuzlu suya, parlatıcıya ve diğer kimyasallara dayanıklıdır.

ELDİVEN ÖLÇÜNÜZÜ ÖĞRENMEK İÇİN;

Elinizi aşağıdaki gibi ölçün. Ölçme parantezimiz size en yakın ölçüde yerleştirilir.



- EN KÜÇÜK
X-SMALL
- KÜÇÜK
SMALL
- ORTA
MEDIUM
- BÜYÜK
LARGE
- EN BÜYÜK
X-LARGE



EN 455-1 & EN 455-2



LEVEL 3 GLOVE



LEVEL 3 GLOVE



Lot No: 12000000000000000000
Çalışma Yeri: 04520-00
Etiler Sok. No: 04520-00
Etiler Sok.
Made in Malaysia

DİKKAT / CAUTION

- Ürünleri ve diğer eldiven diğerleri kullanılmadan kullanılmayın.
- Do not use the product and other gloves product without using the product.

www.nhmedical.com.tr

www.nhmedical.com.tr

